Validation of alternative analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: 3M 01/06-09/97
Renewal decision dated: 29-06-2021
Expiry date: 10-09-2025

The Company:

NEOGEN Food Safety Corporation
620 Lesher Place, Lansing
Michigan 48912, USA

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

Test 3M™ Petrifilm™ *Enterobacteriaceae* Count Plate

Validated for the enumeration of *Enterobacteriaceae*

Technical sheet reference's	**34-8727-3218-4**

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (relative trueness, accuracy profile), available on the certification dedicated website http://nf-validation.org/en.

This certificate supersedes all previous certificates (last version dated 28-07-2021). This NF VALIDATION certificate, included 2 pages, is valid until **September 10th 2025**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).

Managing Director
Julien NIZRI

Issue dated 27/12/2022

11 rue Francis de Pressensé - 93571 La Plaine Saint-Denis Cedex - France - T. +33 (0)1 41 62 80 00 - F. +33 (0)1 49 17 90 00
SAS au capital de 18 187 000 € - 479 076 002 RCS Bobigny - **www.afnor.org**

Validation of alternative analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: 3M 01/06-09/97
Renewal decision dated: 29-06-2021
Expiry date: 10-09-2025

The alternative analysis method:

Test 3M™ Petrifilm™ *Enterobacteriaceae* Count Plate

Validated for the enumeration of *Enterobacteriaceae*

Manufactured by:

3M Wroclaw
ul. Kowalska 143
51-424 Wrocław
POLAND

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	**NF EN ISO 16140-2** (September 2016): Microbiology of the food chain. Method validation - Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method.
Reference method(s)	**NF EN ISO 21528-2** (July 2017) : Microbiology of the food chain - Horizontal method for the detection and enumeration of *Enterobacteriaceae*
Scope	All human food products (by performing validation assays on a broad range of foods), feed products and industrial production environmental samples.
Restriction(s)	None.
Warning	None.
Other informations	Non applicable.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on http://nf-validation.afnor.org/en.

11 rue Francis de Pressensé - 93571 La Plaine Saint-Denis Cedex - France - T. +33 (0)1 41 62 80 00 - F. +33 (0)1 49 17 90 00
SAS au capital de 18 187 000 € - 479 076 002 RCS Bobigny - **www.afnor.org**