

Contacts:

**Mélaïne TERRO**  
Certification Engineer  
[melaine.terro@afnor.org](mailto:melaine.terro@afnor.org)  
Phone: +33 (0)1 41 62 62 39

**Ndatte NDIAYE**  
Operation supervisor  
[ndatte.ndiaye@afnor.org](mailto:ndatte.ndiaye@afnor.org)  
Tél : +33 (0)1 41 62 60 63

Ref.: MET/NND/NF102/Clients/Life Technologies/  
Avis BT\_MicroSEQ L. mono\_2020-01-30\_(R2).docx

Subject: NF VALIDATION mark

**Life Technologies**  
**Part of Thermo Fisher scientific**  
Mrs. Ana-Maria LEONTE  
Wade Road Basingstoke,  
Hampshire RG24 8PW  
United Kingdom

La Plaine Saint-Denis, January 30<sup>th</sup> , 2020

Dear Madam,

Following the positive agreement expressed on 30<sup>th</sup> January 2020, by the Technical Board of the NF VALIDATION mark NF102, in its application to the food industry, I beg to inform you that the **NF VALIDATION certification has been renewed** for the following alternative method:

**MicroSEQ™ *Listeria monocytogenes***

**Certificate reference No. ABI 29/05-12/11, with end of validity 01-Dec.-2023**

The alternative method has been validated by comparison to the reference method NF EN ISO 11290 - 1 (2005) and following the validation protocol NF EN ISO 16140-2 (2016) for the enumeration of *Listeria monocytogenes* in "meat products", "dairy products", "fishery products", "vegetables" and industrial production environmental samples.

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.



Managing Director  
Franck LEBEUGLE

