

Validation of alternative analysis methods

NF102 - Application to the food industry

Certificate

Certificate No.: BIO 12/01-04/94 Renewal decision dated: 17-05-2018

Expiry date: 09-06-2022

The Company:

bioMérieux

Chemin de l'Orme 69280 Marcy l'Etoile - France

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

VIDAS® Salmonella (SLM)

Dual selective enrichment

Validated for the detection of Salmonella spp.

Technical sheet reference's

Ref. 30 702 - 06984 version Y

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The principal certified characteristics are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative detection level ...), available on the certification dedicated website http://nf-validation.afnor.org/en.

This certificate supersedes all previous certificates (last version dated31-01-2018). This NF VALIDATION certificate, included 2 pages, is valid until **09**th **June 2022**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director Franck LEBEUGLE

Page 1/2

Issue dated 27-11-2018









Validation of alternative analysis methods NF102 - Application to the food industry

Certificate

Certificate No.: BIO 12/01-04/94 Renewal decision dated: 17-05-2018

Expiry date: 09-06-2022

The alternative analysis method:

VIDAS® Salmonella (SLM) **Dual selective enrichment**

Validated for the detection of Salmonella spp.

Manufactured by:

bioMérieux Chemin de l'Orme 69280 Marcy l'Etoile - France

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol

NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part

2: Protocol for the validation of alternative (proprietary) methods against a reference method.

Reference method(s)

NF EN ISO 6579-1 (February 2017): Microbiology of the food chain. Horizontal method for the detection, enumeration and serotyping of Salmonella - Part 1: Detection of Salmonella spp.

Scope

All human food products (by performing validation assays on a broad range of foods) and Pet food.

Restriction(s)

None.

Warning

None.

Other informations

The scope includes the use of the following softwares: VIDAS version 4.7, 4.8 and 4.10 and

MiniVIDAS version 5.6.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study summarized report on http://nf-validation.afnor.org/en.





