

Contacts:

**Méline TERRO**  
Certification Engineer  
[melaine.terro@afnor.org](mailto:melaine.terro@afnor.org)  
Phone: +33 (0)1 41 62 62 39

Ref.: MET/MET/NF102/Clients/SOLABIA SAS/  
Avis BT\_Compass Listeria (detection)\_2019-10-03\_(E4).docx

**SOLABIA SAS**  
**Madame Natacha GENTEN**  
2, rue de l'industrie  
60006 BEAUVAIS

Subject: NF VALIDATION mark

La Plaine Saint-Denis, October 7<sup>th</sup>, 2019

Dear Madam,

Following the positive agreement expressed on October 3<sup>rd</sup>, 2019, by the Technical Board of the NF VALIDATION mark (NF102), in its application to the food industry, I beg to inform you that the **extension of NF VALIDATION certification** has been granted to SOLABIA SAS for the following alternative method:

### **Compass® Listeria Agar**

**Certificate reference No. BKR 23/02-11/02, with end of validity 28<sup>th</sup>-Nov.-2023**

(This VALIDATION extension, performed by comparison to the reference method NF EN ISO 11290-1 (2017) and according to the validation protocol NF EN ISO 16140-2 (2016) allowed to validate a new enrichment protocol in Fraser ½ broth at 37°C for the analysis of all human food products (by performing assays on a broad range of food) and industrial production environmental samples.

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.



Managing Director  
Franck LEBEUGLE

