

Contacts:

Méline TERRO
Certification Engineer
melaine.terro@afnor.org
Phone: +33 (0)1 41 62 62 39

Ref.: MET/MET/NF102/Clients/BIO-RAD/
Avis BT_iQ-Check L.mono II_2019-10-04_(E7).docx

BIO-RAD
Mr. Jean-Philippe TOURNIAIRE
3 boulevard Raymond Poincaré
92430 MARNES-LA-COQUETTE

Subject: NF VALIDATION mark

La Plaine Saint-Denis, November 12th, 2019

Dear Sir,

Following the positive agreement expressed on October 4th, 2019, by the Technical Board of the NF VALIDATION mark (NF102), in its application to the food industry, I beg to inform you that the **extension of NF VALIDATION certification** has been granted to BIO-RAD for the following alternative method:


IQ-Check *Listeria monocytogenes* II

Certificate reference No. BRD 07/10-04/05, with end of validity 7th-April-2021

This VALIDATION extension, performed by comparison to the reference method NF EN ISO 11290-1 (2017) and according to the validation protocol NF EN ISO 16140-2 (2016) allowed to extend the scope to the analysis of industrial production environmental samples using the new enrichment specific protocol using LSB broth for 18 - 26h at 30°C, the iQ-Check Free DNA Removal Solution (FDRS) protocol associated with the Easy II lysis protocol and the shortened PCR thermoprofile "Lmono Fast".

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.


Managing Director
Franck LEBEUGLE

