



Validation of alternative
analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **GEN 25/05-11/08**

Extension decision dated: 29-06-2021

Expiry date: 28-11-2024

The Company:

Pall GeneDisc Technologies

1, rue du Courtil
Centre d'affaires CICEA
35170 BRUZ
FRANCE

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

GeneDisc® *Salmonella* spp.

Validated for the detection of *Salmonella* spp.

Technical sheet
reference's

FBIFUGDSLMSPP2ENa

FBIFUGDCAMSLMSSPPENa

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative level of detection), available on the certification dedicated website <http://nf-validation.org/en>.

This certificate supersedes all previous certificates. This NF VALIDATION certificate, included 2 pages, is valid until **November 28th, 2024**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director
Julien NIZRI



Validation of alternative
analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **GEN 25/05-11/08**

Extension decision dated: **29-06-2021**

Expiry date: **28-11-2024**

The alternative analysis method:

GeneDisc® *Salmonella* spp.

Validated for the detection of *Salmonella* spp.

Manufactured by:

Pall GeneDisc Technologies

1, rue du Courtil
Centre d'affaires CICEA
35170 BRUZ
FRANCE

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method.
Reference method(s)	NF EN ISO 6579-1 (April 2017) and its amendment A1 (March 2020): Microbiology of the food chain. Horizontal method for the detection, enumeration and serotyping of <i>Salmonella</i> spp. - Part 1: detection of <i>Salmonella</i> spp.
Scope	All human food products, feed products and industrial production environmental samples.
Restriction(s)	None.
Warning	None.
Other informations	Non applicable.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on <http://nf-validation.afnor.org/en>.