

Contacts:

Stéphanie SAMMARTANO
Business area Manager
stephanie.sammartano@afnor.org
Phone: +33 (0)1 41 62 60 36

Méline TERRO
Certification Engineer
melaine.terro@afnor.org
Phone: +33 (0)1 41 62 62 39

Ref.: SSO/NF148/Clients/IDEXX Laboratories/
Avis BT_Enterolert-DW QT_2021-09-27_R2.docx

Subject: NF VALIDATION mark

IDEXX
Mr. Philippe ROUSSELIN
84 rue Charles Michels
Bâtiment A
F-93200 Saint-Denis - FRANCE

La Plaine Saint-Denis, September 27th, 2021

Dear Sir,

Following the positive agreement expressed on September 27th, 2021, by the Technical Board of the NF VALIDATION mark (NF148), in its application to water analysis, I beg to inform you that the **NF VALIDATION certification is renewed** for the following analysis method:

Enterolert-DW / Quanti-Tray

Certificate reference No. IDX 33/03-10/13, with end of validity 10-Oct.-2025

The alternative method is validated for the enumeration of intestinal enterococci in “drinking waters for human consumption (except bottled waters)”, by comparison to the reference method EN ISO 7899-2 (August 2000) and in accordance with the “validation protocol for an alternative commercial method as compared with a reference method - Revision 2 (17 May 2013)” of the NF VALIDATION certification (NF148).

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.



Stéphanie SAMMARTANO
Business area Manager