

Contacts:

Stéphanie SAMMARTANO
Certification Engineer
stephanie.sammartano@afnor.org
Phone : +33 (0)1 41 62 62 39

Jérôme FAYOL
Operations supervisor
jerome.fayol@afnor.org
Phone : +33 (0)1 41 62 60 63

Ref.: SSO/JFY/NF102/Clients/BioControl Systems/
Avis BT_TAG24 Salmonella_2015-05-22_R2

Subject: NF VALIDATION mark

BioControl Systems
Mr Philip FELDSINE
12822 SE 32nd Street
Bellevue, WA 98005
Etats-Unis

La Plaine Saint-Denis, May 22nd, 2015

Dear Sir,

Following the positive agreement expressed on May 21st, 2015, by the Technical Board "Food microbiology" of the NF VALIDATION mark (NF102), I beg to inform you that the **NF VALIDATION certification is renewed** for the following analysis method:


TRANSIA ADDITIVE *Salmonella* Gold 24h

Certificate reference No. TRA 02/09-07/07, with end of validity 2nd-July-2019

The method is validated for the detection of *Salmonella* spp. in all human and animal food products, by comparison to the reference method NF EN ISO 6579 :2002 and following the validation protocol NF EN ISO 16140 :2003.

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely,


Managing Director
Franck LEBEUGLE



Contact :

Stéphanie SAMMARTANO
stephanie.sammartano@afnor.org
Phone : +33 (0)1 41 62 62 39
Fax : + 33 (0)1 49 17 90 86

Ref.: SSO/NF102/Clients/ BioControl Systems/
Avis BT décembre 2011 Reconduction TAG 24 Salmonella

BIOCONTROL SYSTEMS
Monsieur Christophe BAUER
37, Gratteloup
45490 LORCY
France

Subject: NF VALIDATION

Saint-Denis, December 5th, 2011

Dear Sir,

Following the positive agreement expressed on December 1st, 2011, by the NF VALIDATION Technical Committee, in its application to agribusiness, I beg to inform you that the NF VALIDATION certification is renewed for the following analysis alternative method:

TAG 24 *Salmonella*

For the detection of *Salmonella spp* in human food and animal feeding stuffs

Certificate reference n# TRA 02/09 – 07/07, with end of validity : 2-July-2015

A further letter will mention full conclusions and possible reservations made by the Technical Committee. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.



Managing Director
Florence MÉAUX

