

Contacts:

Méline TERRO  
Certification Engineer  
[melaine.terro@afnor.org](mailto:melaine.terro@afnor.org)  
Phone: +33 (0)1 41 62 62 39

Ref.: MET/MET/NF102/Clients/Solus Scientific Solutions/  
Avis BT\_Solus E. coli O157\_2019-10-03\_(R1).docx

**Solus Scientific Solutions, Ltd.**  
**Mr. Simon ILLINGWORTH**  
Unit 9 Mansfield Networkcentre  
Millennium Business Park  
Concorde Way  
NG18 5BR Mansfield (UK)

Subject: NF VALIDATION mark

La Plaine Saint-Denis, October 7<sup>th</sup>, 2019

Dear Sir,

Following the positive agreement expressed on October 3<sup>rd</sup>, 2019, by the Technical Board of the NF VALIDATION mark (NF102), in its application to the food industry, I beg to inform you that the **NF VALIDATION certification has been renewed** for the following alternative method:

### **Solus *E. coli* O157 ELISA**

**Certificate reference No.SOL 37/03-10/15, with end of validity 14<sup>th</sup>-Oct.-2023**

The alternative method has been validated by comparison to the reference method NF EN ISO 16654/A1 (2017) and following the validation protocol NF EN ISO 16140-2 (2016) for the detection of *Escherichia coli* O157 (including E.coli O157:H7) in raw beef meat products (seasoned or not), raw milk and dairy products, vegetables products and industrial production environmental samples.

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely,



Managing Director  
Franck LEBEUGLE

