

Contacts:

Stéphanie SAMMARTANO
Senior Certification Engineer
stephanie.sammartano@afnor.org
Phone: +33 (0)1 41 62 62 39

Jérôme FAYOL
Operations supervisor
jerome.fayol@afnor.org
Phone: +33 (0)1 41 62 60 63

Ref.: SSO/JFY/NF102/Clients/ Prestodiag/
Avis BT_SalmoPresto PE_2017-07-03_(V).docx

Subject: NF VALIDATION mark

Prestodiag SAS
Mr. Felix PIAT
Villejuif Bio Park
1 mail du professeur Georges Mathé
F-94800 VILLEJUIF, France

La Plaine Saint-Denis, July 3rd, 2017

Dear Sir,

Following the positive agreement expressed on July 3rd, 2017, by the Technical Board of the NF VALIDATION mark (NF102), in its application to the "food industry", I beg to inform you that the **NF VALIDATION certification has been granted** to Prestodiag for the following alternative method:


***SalmoPresto*TM PE**

Certificate reference No. PDG 39/01-07/17, with end of validity 3rd-July-2021

The alternative method has been validated for the detection of *Salmonella* spp. in egg products, by comparison to the reference method NF EN ISO 6579 (2002) and following the validation protocol NF EN ISO 16140-2 (2016). The validated scope includes the use of the MonoPrestoTM reader with use with the software Bacterics version 1.1.2.2.

A further letter will mention full conclusions and possible reservations made by the Technical Board. If reservations are mentioned, I ask you to take them into account without any delay.

Yours Sincerely.



Managing Director
Franck LEBEUGLE

