

TO REMIND IN ALL CORRESPONDANCE

No. File: MET/MET/NF102/Clients/IDEXX Laboratories/
Avis BT_IDEXX SNAP Duo ST Plus_2020-10-02_(V).docx
No. Application: NF102

DECISION RELATED TO THE USAGE RIGHT

File followed by the Department: **PGCP6**
Certification Engineer: **Méline TERRO**
Phone: +33 (0)1 41 62 62 39
Email: melaine.terro@afnor.org

IDEXX Laboratories, Inc.
One Drive
Westbrook,
Maine 04092, USA

To the attn. of **Mr. Jean-Luc TROCH**

Subject: NF VALIDATION mark

La Plaine Saint-Denis, October 2nd, 2020

Dear Sir,

Following the positive agreement expressed on October 2nd, 2020, by the Technical Board of the NF VALIDATION mark (NF102), in its application to the food industry, I beg to inform you that the **NF VALIDATION certification has been granted** to IDEXX Laboratories, Inc. for the following alternative method:

IDEXX SNAPduo ST Plus Test

Certificate reference No. IDX 33/08-10/20, with end of validity 2nd-Oct.-2024

The alternative method has been validated for the detection of beta-lactams residues (penicillins and cephalosporins including cefalexin) and tetracycline in raw commingled and individual cow milk with visual and instrumental reading, according to the validation protocol for detection methods for residues of veterinary drugs in food matrices - Rev. 1 (2017) of the NF VALIDATION mark, in its application to the food industry (NF102).

Yours Sincerely.



Méline TERRO
Certification Engineer