

Certificate

Certificate No.: BRD 07/10-04/05 Extension decision dated: 04-10-2023

Expiry date: 07-04-2025

NF102 - Application to the food industry

The Company:

BIO-RAD

3 Boulevard Raymond Poincaré 92430 Marnes La Coquette - France

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

iQ-Check® Listeria monocytogenes II

Validated for the detection of *Listeria monocytogenes*

Technical sheet reference's

Cat. 10000167776 Ver A

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The principal certified characteristics are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative detection level...), available on the certification dedicated website http://nf-validation.afnor.org/en.

This certificate supersedes all previous certificates (last version dated 17-04-2023). This NF VALIDATION certificate, included 2 pages, is valid until 07th April 2025. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director Julien NIZRI

Issue dated 24/10/2023



Certificate

Certificate No.: BRD 07/10-04/05 Extension decision dated: 04-10-2023

Expiry date: 07-04-2025

NF102 - Application to the food industry

The alternative analysis method:

iQ-Check® Listeria monocytogenes II

Validated for the detection of *Listeria monocytogenes*

Manufactured by:

BIO-RAD Laboratories, Inc 2000 Alfred Nobel Drive California 94547 Hercules - USA

Has been certified according to the reference documents and for the application scope specified here after:

NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part 2: Validation protocol

Protocol for the validation of alternative (proprietary) methods against a reference method.

Reference method(s) NF EN ISO 11290-1 (July 2017): Microbiology of the food chain - Horizontal method for the

detection and enumeration of Listeria monocytogenes - Part 1: detection method.

All human products (by performing validation assays on a broad range of foods) and industrial Scope

production environment samples.

Restriction(s) None.

> Warning None.

Other information(s) Scope includes use of CFX96 Touch, CFX96 Touch Deep well and CFX OPUS Deepwell real-

time PCR instruments with CFX Manager IDE v3.1 software

Scope includes use of a new enrichment environment of Listeria Special Broth II (LSB II) for

dairy products, composite foods and industrial production environmental samples.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study summarized report on http://nf-validation.afnor.org/en.

