



Validation of alternative  
analysis methods

NF102 – Application to the food industry

# Certificate

Certificat No.: **3M 01/02-09/89 C**

Renewal decision dated: **08-06-2022**

Expiry date: **09-06-2026**

The Company:

**NEOGEN Food Safety Corporation**  
620 Lesher Place, Lansing  
Michigan 48912, USA

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

## Neogen®Petrifilm® Coliform Count (CC) Plate

Validated for the enumeration of thermotolerant coliforms

Technical sheet  
reference's

FS00931A

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (linearity, relative accuracy...), available on the certification dedicated website <http://nf-validation.org/en>.

This certificate supersedes all previous certificates (last version dated 23-12-2022). This NF VALIDATION certificate, included 2 pages, is valid until **9<sup>th</sup> June 2026**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



  
Managing Director  
Julien NIZRI

Issue dated 03-01-2024

Page 1/2



Validation of alternative  
analysis methods

NF102 – Application to the food industry

# Certificate

Certificat No.: **3M 01/02-09/89 C**

Renewal decision dated: **08-06-2022**

Expiry date: **09-06-2026**

The alternative analysis method:

## Neogen® Petrifilm® Coliform Count (CC) Plate

Validated for the enumeration of thermotolerant coliforms

Manufactured by:

**3M Wrocław**  
ul. Kowalska 143  
51-424 Wrocław  
POLAND

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	<b>NF EN ISO 16140-2</b> (September 2016) Microbiology of the food chain - Method validation - Part 2: protocol for the validation of alternative (proprietary) methods against a reference method.
Reference method(s)	<b>NF V08-060</b> (April 2009): Microbiology of food and animal feeding stuffs. Enumeration of thermotolerant <i>coliforms</i> by colony-count technique at 44 °C.
Scope	All human food products (by performing validation assays on a broad range of foods).
Restriction(s)	None.
Warning	None.
Other informations	Non applicable.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on <http://nf-validation.afnor.org/en>.

CERTIF 2022.1 08/2020

Issue dated 03-01-2024

Page 2/2