



Validation of alternative analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **BIO 12/32-10/11**

Renewal decision dated: **15-06-2023**

Expiry date: **06-10-2027**

The Company:

bioMérieux

Chemin de l'Orme

69280 Marcy l'Etoile - France

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

VIDAS[®] Up *Salmonella* (SPT)

Validated for the detection of *Salmonella* spp.

Technical sheet reference's

Ref. 30 707 – 044875 version **06**

Ref. 30 707 – 044875 version **07**

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative detection level ...), available on the certification dedicated website <http://nf-validation.afnor.org/en>.

This certificate supersedes all previous certificates (last version dated 07-06-2024). This NF VALIDATION certificate, included 2 pages, is valid until **October 6th, 2027**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director
Julien NIZRI

Issue dated 06/11/2024

Page 1/2



Validation of alternative analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **BIO 12/32-10/11**

Renewal decision dated: **15-06-2023**

Expiry date: **06-10-2027**

The alternative analysis method:

VIDAS[®] Up *Salmonella* (SPT)

Validated for the detection of *Salmonella* spp.

Manufactured by:

bioMérieux
Chemin de l'Orme
69280 Marcy l'Etoile - France

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method.
Reference method(s)	NF EN ISO 6579-1 (April 2017) and its amendment A1 (March 2020): Microbiology of the food chain. Horizontal method for the detection, enumeration and serotyping of <i>Salmonella</i> spp. - Part 1: detection of <i>Salmonella</i> spp. - Amendment 1 Broader range of incubation temperatures, amendment to the status of Annex D, and correction of the composition of MSR/V and SC.
Scope	All human food products (by performing validation assays on a broad range of foods), pet food products and animal feed products, industrial and primary production environmental samples.
Restriction(s)	None.
Warning	None.
Other informations	The scope includes the use of the following softwares: VIDAS version 4.10 and previous, MiniVIDAS version 5.6, VIDAS KUBE version 1.1 and previous, VITEK [®] MS version 3.0 and previous, VITEK [®] MS Prime.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on <http://nf-validation.afnor.org/en>.