



Validation of alternative
analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **BRD 07/01-07/93**

Extension decision dated: **15-10-2025**

Expiry date: **02-12-2028**

The Company:

BIO-RAD

3 Bd Raymond Poincaré
92430 Marnes La Coquette FRANCE

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

RAPID' *E. coli* 2 (*E. coli* 44°C)

Validated for the enumeration at 44°C of β -glucuronidase positive *Escherichia coli*

Technical sheet
reference's

10000243822 **Ver B**

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (relative trueness, accuracy profile), available on the certification dedicated website <http://nf-validation.org/en>.

This certificate supersedes all previous certificates (last version dated 04-10-2024). This NF VALIDATION certificate, included 2 pages, is valid until **December 2nd 2028**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director
Julien NIZRI

Issue dated 15/10/2025

Page 1/2



Validation of alternative
analysis methods

NF102 – Application to the food industry

Certificate

Certificate No.: **BRD 07/01-07/93**

Extension decision dated: **15-10-2025**

Expiry date: **02-12-2028**

The alternative analysis method:

RAPID' *E. coli* 2 (*E. coli* 44°C)

Validated for the enumeration at 44°C of β -glucuronidase positive *Escherichia coli*

Manufactured by:

BIO-RAD

Route de Cassel

59114 STEENVOORDE FRANCE

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method.
Reference method(s)	NF ISO 16649-2 (July 2001) : Microbiology of food and animal feeding stuffs - Horizontal method for the enumeration of B-glucuronidase-positive <i>Escherichia coli</i> - Part 2 : colony-count technique at 44 °C using 5-bromo-4-chloro-3-indolyl B-D-glucuronate
Scope	All human food products (by performing validation assays on a broad range of foods).
Restriction(s)	None.
Warning	None.
Other informations	The scope includes the option of using a single plate and a single dilution (refer to the method's instructions for use).

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on <http://nf-validation.afnor.org/en>.

CERTI F 2022.1 08/2020

Issue dated 15/10/2025

Page 2/2