



Validation of analysis methods

NF148 – Application to water

Certificate

Certificate No.: **BRD 07/19-11/09**

Renewal decision dated: **08-10-2025**

Expiry date: **06-11-2029**

The Company:

BIO-RAD

3 Boulevard Raymond Poincaré
92430 Marnes La Coquette - France

Is authorized to affix the NF VALIDATION mark on the alternative method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF148 - Validation of analysis methods (Application to water):

XplOrer64™ / CheckN'Safe™ *Enterococci*

Validated for the enumeration of intestinal *enterococci*

Technical sheet
reference's

CheckN'Safe™ *Enterococci* test (Code 3554721) : **V3.5**

XplOrer64™ System User Manual: **V3.0**

This decision attests that the analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (linearity, sensitivity, relative detection level...), available on the certification dedicated website <http://nf-validation.org/en>.

This certificate supersedes all previous certificates (last issue dated 18-07-2024). This NF VALIDATION certificate, included 2 pages, is valid until **November 06th, 2029**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF148 - Validation of alternative analysis methods (Application to water).



Managing Director
Julien NIZRI



Validation of analysis methods
NF148 – Application to water

Certificate

Certificate No.: **BRD 07/19-11/09**
Renewal decision dated: **08-10-2025**
Expiry date: **06-11-2029**

The alternative analysis method:

XplOrer64™ / CheckN'Safe™ *Enterococci*

Validated for the enumeration of intestinal *enterococci*

Manufactured by:

SY-LAB Geräte GmbH
Tullnerbachstrasse 61-65
3011 Neupurkersdorf - Austria

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol	NF VALIDATION - Validation of analysis methods: Application to water microbiology. Validation protocol for an alternative commercial method as compared with a reference method. Revision 2 (May 2013).
Reference method(s)	NF EN ISO 7899-1 (March 1999): Water quality. Detection and enumeration of intestinal <i>enterococci</i> in surface and waste water - Part 1: miniaturized method (Most Probable Number) by inoculation in liquid medium.
Scope	Bathing waters (sea water and fresh water); waste waters, raw and processed waste waters
Restriction(s)	None.
Warning	None.
Other information	The scope includes the use of the XplOrer64™ automated System with use with XplOrer64™ software V3.0.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study **summarized report** on <http://nf-validation.afnor.org/en>.