

Validation of alternative analysis methods

NF102 - Application to the food industry

Certificate

Certificate No.: BKR 23/02-11/02 Renewal decision dated: 17-10-2023

Expiry date: 28-11-2027

The Company:

SOLABIA

41, Rue Delizy 93698 Pantin cedex FRANCE

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

COMPASS® Listeria Agar (detection)

Validated for the detection of Listeria spp. and Listeria monocytogenes

Technical sheet reference's

COMPASS LISTERIA_FR_V17

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative level of detection...) available on the certification dedicated website http://nf-validation.org/en.

This certificate supersedes all previous certificates (last version dated 25-10-2021). This NF VALIDATION certificate, included 2 pages, is valid until **November 28**th, **2027**. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director
Julien NIZRI

Issue dated 20/10/2023

Page 1/2





Certificate No.: BKR 23/02-11/02 Renewal decision dated: 17-10-2023

Expiry date: 28-11-2027

The alternative analysis method:

COMPASS® Listeria Agar (détection)

Validated for the detection of Listeria spp. and Listeria monocytogenes

Manufactured by:

SOLABIA

ZAC de Ther –Allonne Rue des 40 mines – BP 10245 6002 Beauvais Cedex FRANCE

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol NF EN ISO 16140-2 (September 2016): Microbiology of the food chain. Method validation - Part

2: Protocol for the validation of alternative (proprietary) methods against a reference method.

Reference method(s) NF EN ISO 11290-1 (July 2017): Microbiology of the food chain - Horizontal method for the

detection and enumeration of Listeria monocytogenes and of Listeria spp. - Part 1: detection

method.

Scope All human food products (by performing validation assays on a broad range of foods) and industrial

production environmental samples.

Restriction(s) None.

Warning None.

Other information Non applicable.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study summarized report on http://nf-validation.afnor.org/en.

